RoHS 2011/65/EU & (EU) 2015/863 COMPLIANT

CUSTOMER:	P/N: 0313.240BAL	REV. .	APPROVED BY:

REVISIONS		
REV.	DESCRIPTION	DATE
A	The plug insert changed to the exposed type,Update connector mark	2024/07/09
B	Update cable safety regulation and connector mark The plug insert changed to the exposed type Change label specifications and content, add note 2	2024/12/02
C	Add Lot annotation(use Customer PO)	2026/01/13



NOTES:

1. RoHS shown on the carton
2. connector SR apply molded type

MARKING:



NEMA 5-15P
HOSPITAL GRADE



MARKING:



IEC-60320-C13
AUTO-LOCK

⑤ Hosp.Grade label(English/French version+Customer part number)



ITEM	DESCRIPTION	COLOR	QTY
⑥	Cable tie	BLACK	2
⑤	Hosp.Grade label(English/French version+Customer part number)	---	1
④	PLASTIC CAP	WHITE	1
③	SJTW 14AWG 3C 105°C APPROVAL: UL,CUL	BLACK	1
②	WS-002A-4A 15(A) 125(V) APPROVAL: UL,CUL	CLEAR	1
①	WS-001H 15(A) 125(V) APPROVAL: UL,CUL	CLEAR	1

	SIGNATURE
ENGINEER BY:	
MANAGER BY:	
CHECKED BY:	
DRAWN BY:	

QUAIL ELECTRONICS, INC. ®
http://www.quail.com

PRODUCT NO.:	DATE: 2026/01/13	
UNIT: mm	SHEET: 1 OF 1	REV.
DWG NO: 0096-6075		C

THE INFORMATION CONTAINED HEREIN IS PROPRIETARY TO QUAIL ELECTRONICS, INC. ANY USE OR DISCLOSURE OF THIS INFORMATION, EXCEPT FOR THE EXCLUSIVE BENEFIT OF QUAIL ELECTRONICS, INC. IS PROHIBITED.