RoHS (2011/65/EU) COMPLIANT

CUSTOMER:	P/N: 0322.144 REV.	APPROVED BY:

REVISIONS

REV.	DESCRIPTION	DATE
A		2010.01.26
B		2010.02.22
C		2013/11/05
D	Update head print and note 1.2, modify label content	2017/04/06





NEMA 5-15P HOSPITAL GRADE CLEAR



MARKING:

WELL SHIN Ⓡ
WS-002
13A 125V~
E115330



IEC-60320-C13 CLEAR

NOTES:

1. The cable has CSA on ZH(Zhang Hwa) factory ZH(Zhang Hwa) factory begin from revision C
2. RoHS shown on the carton

⑥ Hosp.Grade label



ITEM	DESCRIPTION	COLOR	Q'TY
⑥	Hosp.Grade label	---	1
⑤	Cable tie	WHITE	1
④	PLASTIC CAP	WHITE	1
③	SJT 16AWG 3C 105°C APPROVAL : UL,C-UL(CSA)	Pantone Cool Grey 9C (GY-H)	1
②	WS-002 13 (A) 125(V) APPROVAL : UL,C-UL	CLEAR	1
①	WS-001H 13 (A) 125(V) APPROVAL : UL,C-UL	CLEAR	1

	SIGNATURE
SAFETY ENGINEER:	
MANAGER BY:	
CHECKED BY:	
DRAWN BY:	

QUAIL ELECTRONICS, INC.
http://www.quail.com

PRODUCT NO.: 001H+002-1	DATE: 2017/04/06	
UNIT: mm	SHEET: 1 OF 1	REV
DWG NO: 0096-0392		D

THE INFORMATION CONTAINED HEREIN IS PROPRIETARY TO QUAIL ELECTRONICS, INC. ANY USE OR DISCLOSURE OF THIS INFORMATION, EXCEPT FOR THE EXCLUSIVE BENEFIT OF QUAIL ELECTRONICS, INC. IS PROHIBITED.

Copyright 2017 Quail Electronics